

Certificate No.: NA0020-269-011117

Mettler Toledo
Service Business Unit Industrial
1900 Polaris Parkway
Columbus, OH 43240

METTLER TOLEDO

ISO 9001: 2008 Registered

Basic Test Confirmation

Customer

Company:	Smith Co.		
Address:	100 Smith Way		
City:	Anywhere	State/Province:	Pa
Zip/Postal:	00000		
Terminal Type:	Wall Mt	Serial No. Terminal:	00000-00
Procedure Statement:	The device referenced in this document has been metrologically tested in accordance with METTLER TOLEDO Work Instruction VW0152A. All translations into other languages are based on the referenced work instruction, which is in English. This certificate refers to: As Found and As Left		
Test Date:	11-Jan-2017	Next Cal. Due Date:	11-May-2017
Service Technician:	Frank Gardone	Signature:	ELECTRONIC SIGNATURE

Reference Weights

Traceability of Test Equipment:	All weights used for metrological testing are traceable to national or international standards. The weights were calibrated and certified by an accredited calibration laboratory.
---------------------------------	--

Weight Set 1

Weight Set No.:	156	Date of Issue:	26-Aug-2013
Calibration Due Date:	24-Aug-2016	Certificate Number:	66666
Class:	NIST F		

Weight Set 2

Weight Set No.:	100-120	Date of Issue:	31-Aug-2013
Calibration Due Date:	2-Feb-2017	Certificate Number:	555
Class:	NIST F		

List of Devices Tested

<i>Manufacturer</i>	<i>Model</i>	<i>Serial No.</i>	<i>Ecc</i>	<i>Lin</i>	<i>Hys</i>	<i>Weight Set</i>	<i>Ref. Wt. Span</i>	<i>AS FOUND</i>	<i>AS LEFT</i>
Mettler Toledo	2181	011				156	1000 lb	1000 lb	1000 lb
Mettler Toledo	2181	112	✓	✓		156	1000 lb	1005 lb	1000 lb
Mettler Toledo	Mettler Toledo	113	✓	✓	✓	156	500 lb	500 lb	500 lb
Mettler Toledo	2181	114	✓	✓	✓	156	500 lb	500 lb	500 lb

Remarks

Area Clean and dry. No damage or anomalies noted on arrival.