## Properties of Solutions: Electrolytes and Non-Electrolytes

- 1. Editable Microsoft Word versions of the student pages and pre-configured TI-Nspire files can be found on the CD that accompanies this book. See *Appendix A* for more information.
- 2. We suggest that you set up the Conductivity Probes before the experiment. Set the selection switch on the amplifier box of the probe to the  $0-20000 \ \mu\text{S/cm}$  range.
- 3. Fewer sets of Groups A, B, and C can be prepared if students are advised that they need not start with Group A. Add solutions to 100 mL beakers or small vials to a depth that easily allows the hole near the Conductivity Probe tip to be completely submerged (the graphite electrodes of the probe are located on either side of this hole).
- 4. Preparation of solutions (prepare all solutions in distilled water):

0.050 M CaCl<sub>2</sub> (5.55 g of solid calcium chloride, CaCl<sub>2</sub>, per 1 L solution) Hazard Code: D—Relatively non-hazardous. Alternatively, 7.35 g CaCl<sub>2</sub>•2H<sub>2</sub>O, per 1 L solution. **HAZARD ALERT:** Toxic by ingestion. Hazard Code: D—Relatively non-hazardous.

0.050 M NaCl (2.93 g of solid sodium chloride, NaCl, per 1 L solution) **HAZARD ALERT:** Moderately toxic. Hazard Code: D—Relatively non-hazardous.

0.050 M AlCl<sub>3</sub> (12.05 g of solid aluminum chloride, AlCl<sub>3</sub>•6H<sub>2</sub>O, per 1 L solution) preferred. Hazard Code: D—Relatively non-hazardous. Alternatively, 6.67 g anhydrous AlCl<sub>3</sub> per liter of solution. **HAZARD ALERT:** Reacts very violently with water; toxic by inhalation and ingestion; strong skin irritant. Hazard Code: A—Extremely hazardous.

0.050 M HCl (4.2 mL of concentrated hydrochloric acid, HCl, per 1 L solution) **HAZARD ALERT:** Highly toxic by ingestion or inhalation; severely corrosive to skin and eyes. Hazard Code: A—Extremely hazardous.

0.050 M HC<sub>2</sub>H<sub>3</sub>O<sub>2</sub> (2.9 mL of concentrated acetic acid, HC<sub>2</sub>H<sub>3</sub>O<sub>2</sub>, per 1 L solution) **HAZARD ALERT:** Corrosive to skin and tissue; moderate fire risk (flash point: 39°C); moderately toxic by ingestion and inhalation. Hazard Code: A—Extremely hazardous.

0.050 M H<sub>3</sub>PO<sub>4</sub> (3.4 mL of concentrated phosphoric acid, H<sub>3</sub>PO<sub>4</sub>, per 1 L solution) **HAZARD ALERT:** Skin and eye irritant; moderately toxic by ingestion and inhalation; corrosive; burns tissue. Hazard Code: A—Extremely hazardous.

0.050 M H<sub>3</sub>BO<sub>3</sub> (3.09 g of solid boric acid, H<sub>3</sub>BO<sub>3</sub>, per 1 L solution) **HAZARD ALERT:** Moderately toxic by ingestion; irritant to skin in dry form. Hazard Code: C—Somewhat hazardous.

0.050 M CH<sub>3</sub>OH (1.60 g (2.1 mL) methanol per 1 L solution) **HAZARD ALERT:** Flammable; dangerous fire risk; toxic by ingestion (ingestion may cause blindness). Hazard Code: B—Hazardous.

The hazard information reference is: Flinn Scientific, Inc., *Chemical & Biological Catalog Reference Manual*, (800) 452-1261, www.flinnsci.com. See *Appendix F* for more information.

- 5. Conductivity readings are normally reported in microsiemens per centimeter, or  $\mu$ S/cm. This SI derived unit has replaced the conductivity unit, micromho/cm.
- 6. Students are instructed to rinse the probe with distilled water between samples. They are told to blot the probe tip dry—however, the directions also remind them that they do *not* need to blot dry the inside of the hole containing the graphite electrodes. It is cumbersome to do so, and leaving a drop or two of distilled water does not significantly dilute the next sample.
- 7. Using the stored calibration, measured conductivity values for  $H_3BO_3$ ,  $CH_3OH$ , or distilled water will be in the range of 0 to 30  $\mu$ S/cm. If a two-point calibration is performed, students will get readings closer to 0  $\mu$ S/cm. These four samples will usually have a small conductivity value due to dissolved carbon dioxide, which forms aqueous ions according to the equation:

 $CO_2(g) + H_2O(l) \longleftrightarrow H^+(aq) + HCO_3^-(aq)$ 

The resulting conductivity, usually about  $1-3 \mu$ S/cm, can be accurately measured using the narrower 0–200  $\mu$ S/cm setting and calibration for the Conductivity Probe. You could do this as a teacher demonstration, or instruct your students to do it as an extension to the experiment.

At the 0–200  $\mu$ S/cm setting, students will also notice that the conductivity of boric acid is higher than distilled water, 0.05 M methanol, or 0.05 M ethylene glycol. This way, they can see that boric acid is a weak acid that ionizes to a very small extent. For example, we get a reading of 3.2  $\mu$ S/cm for 0.05 M boric acid, but only 1.0  $\mu$ S/cm for distilled water, and 1.0  $\mu$ S/cm for 0.05 M methanol, using the 0–200  $\mu$ S/cm setting.

8. If you wish to calibrate the Conductivity Probe to improve conductivity readings at low concentrations (as discussed in item 7 above), follow these directions:

First Calibration Point

- a. Choose Set Up Sensors ► Calibrate ► Two Point from the 🖉 Experiment menu.
- b. For the first calibration point, the Conductivity Probe should simply be in the air (out of any liquid or solution).
- c. Enter **0** as the first reference value.
- d. When the voltage stabilizes, select OK.

Second Calibration Point

- e. Place the Conductivity Probe into a standard solution that is equivalent to 10,000  $\mu$ S/cm. **Note:** This standard can be prepared by dissolving 5.566 g of solid sodium chloride, NaCl, in enough distilled water for 1 liter of solution.
- f. Enter 10000 as the second reference value (in  $\mu$ S/cm).
- g. When the voltage stabilizes, select OK.
- h. Select OK to complete the calibration.

## SAMPLE RESULTS

Solution	Conductivity (µS/cm)
A - CaCl <sub>2</sub>	9362
A - NaCl	5214
A - AICI <sub>3</sub>	11707
B - HC <sub>2</sub> H <sub>3</sub> O <sub>2</sub>	461
B - HCI	17330
B - H <sub>3</sub> PO <sub>4</sub>	6661
B - H <sub>3</sub> BO <sub>3</sub>	0
C - H <sub>2</sub> Odistilled	0
C - H <sub>2</sub> O <sub>tap</sub>	(varies) 20 – 1000
C - CH₃OH	0

## **ANSWERS TO QUESTIONS**

For Sample Answers to the questions in this lab, please contact Vernier Software and Technology at <u>swnanswers@vernier.com</u>